



Purchasing Services Office

801 Leroy Place
Socorro, NM 87801
(575) 835-5886

Letter of Addendum

TO: All Offerors

FROM: Meradeth Montoya, Associate Director of Purchasing

DATE: 3/30/2026

RE: RFP Number: RFP26030025- Amendment No. 1
Commodity: Wavelength Dispersive X-ray Fluorescence Spectrometer

- Q1. In Section 15.2, the last bullet point states that instrument dimensions shall not exceed Width < 40 inches, Depth < 40 inches, and Height < 60 inches. Since the depth exceeds the stated limit, does this disqualify us from the bid?
- A1. This is not disqualifying, provided the width remains within the maximum requirement. However, overall instrument size will be taken into consideration to ensure adequate working space is maintained within the laboratory.

ALL OFFERORS ARE REQUIRED TO CONFIRM THE RECEIPT OF THIS AMENDMENT IN THEIR RESPONSE.
ALL OTHER TERMS AND CONDITIONS OF THE RFP REMAIN UNCHANGED.