

***Letter of Addendum***

TO: All Offerors

FROM: Meradeth Montoya, Associate Director of Purchasing

DATE: 05/08/2026

RE: RFP Number: RFP 26050031 – Amendment No. 1  
Commodity: Rock Triaxial Testing System

Q1: Section 5.1 Project Overview

Control and Wireless Servo Controller and Data Acquisition System, Is this required or recommended? Wireless servos have a higher chance of interference and data loss.

A1: Wireless servo control is preferred but not required.

Q2: Section 15.2 Technical Requirements and Specifications, Contrary specifications are listed. Which specification is correct?

- Axial load capacity  $\geq$  500 kN

A2: The correct specification is: Axial load capacity  $\geq$  500 kN.

Q3: Section 15.5.1 Current Description of Installation Location, Power available: 250 Volt, 30 Amp single phase. Most hydraulic power supplies (Hydraulic Power Unit: 10HP and 5GPM) require 3 phase power. Is this acceptable?

A3: Yes, 3 phase power is acceptable.

ALL OFFERORS ARE REQUIRED TO CONFIRM THE RECEIPT OF THIS AMENDMENT IN THEIR RESPONSE.  
ALL OTHER TERMS AND CONDITIONS OF THE RFP REMAIN UNCHANGED.