



Purchasing Services Office

801 Leroy Place
Socorro, NM 87801
(575) 835-5886

Letter of Addendum

TO: All Offerors

FROM: Lisa Majkowski, Director – Purchasing Services

DATE: 09/18/2025

RE: RFB Number: RFP 26090007 - Amendment No. 2
Commodity: Room Calorimeter

Please note the following corrections and clarifications (**in red**) to the equipment specifications:

PART 1 – General

2. SUMMARY

A. Section Includes

1. Furniture Calorimeter / Room size calorimeter (without the room) / Large Scale Heat Release Apparatus
 - a. The objectives are **operate at a minimum of 100 kW for 10 minutes and then 300 kW for 10 min in accordance with ISO 9705.**

PART 2 – Products

1. Unit Requirements

B. Sensors (all sensor must be able to perform function in system)

1. Gas Analysis Instrumentation, Oxygen (0-25%), Carbon Dioxide (0-10%), Carbon Monoxide (0-1%), and Humidity (**±2% RH**).
3. Flow Rate: Bi-directional **probe** with K-type thermocouple capable of KT-values in accordance with ASTM 603.

ALL OFFERORS ARE REQUIRED TO CONFIRM THE RECEIPT OF THIS AMENDMENT IN THEIR RESPONSE.
ALL OTHER TERMS AND CONDITIONS OF THE RFP REMAIN UNCHANGED.