



Purchasing Services Office

801 Leroy Place
Socorro, NM 87801
(575) 835-5886

Letter of Addendum

TO: All Offerors

FROM: Meradeth Montoya, Associate Director of Purchasing

DATE: 04/13/2026

RE: RFP Number: RFP26030025- Amendment No. 3
Commodity: Wavelength Dispersive X-ray Fluorescence Spectrometer

Q1: One question regarding the dimensions outlined in Section 15.2 Equipment Description and Specification on Page 11 of the RFP. The requirement states: "Instrument dimensions shall not exceed the following: Width less than 40 inches, Depth less than 40 inches, Height less than 60 inches."

Our instrument dimensions are 44.1" (W) x 33.1" (D) x 59.5" (H) with the sample changer cover closed, and 86.2" in height when fully open. We would like to note that there will be no issue with accessing the laboratory, as the system is equipped with wheels that allow it to be maneuvered through the doorway on its narrower side.

Would you please confirm whether these dimensions are acceptable?

A1: Yes, these dimensions are acceptable.

ALL OFFERORS ARE REQUIRED TO CONFIRM THE RECEIPT OF THIS AMENDMENT IN THEIR RESPONSE.
ALL OTHER TERMS AND CONDITIONS OF THE RFP REMAIN UNCHANGED.